



INSTYTUT CHEMII I TECHNIKI JĄDROWEJ

POLISH CERTIFIED REFERENCE MATERIAL

FOR MULTIELEMENT TRACE ANALYSIS WITH THE SPECIAL EMPHASIS ON RARE EARTH ELEMENTS

APATITE CONCENTRATE (CTA-AC-1)

It has been complied
with the requirements of ISO Guide 34.

GENERAL INFORMATION

INTENDED USE:

CHECKING THE ACCURACY OF ANALYTICAL WORK OF THE LABORATORIES ENGAGED IN THE DETERMINATION OF TRACE ELEMENTS IN MINERAL SAMPLES. CALIBRATION OF APPARATUS AND METHODS.

ELEMENTS FOR WHICH CERTIFIED CONCENTRATIONS COULD BE ESTABLISHED:

Ba, Ca, Ce, Co, Cu, Eu, Gd, Hf, La, Lu, Mn, Na, Nd, Sc, Si, Sm, Ta, Tb, Th, Ti, U, V, Y, Yb, Zn.

ELEMENTS FOR WHICH NON-CERTIFIED "INFORMATION" VALUES ARE AVAILABLE:

Al, Dy, Er, Fe, Ho, K, Mg, Ni, Pr, Sr, Zr.

Produced and certified by:

Distributed by:

COMMISSION OF TRACE ANALYSIS OF THE COMMITTEE FOR ANALYTICAL CHEMISTRY OF THE POLISH ACADEMY OF SCIENCES

Zakład Chemii Analitycznej ICHiTJ
ul. Dorodna 16
03-195 Warszawa, Poland

and

INSTITUTE OF NUCLEAR CHEMISTRY AND TECHNOLOGY, UL. DORODNA 16, 03-195 WARSZAWA, POLAND

Phone: 112737

Telex: 813027 ichtj pl.

ORIGIN, PREPARATION AND TESTING

Apatite concentrate (ca. 50 kg) originating from Kola Peninsula (USSR), and used in Polish chemical industry as a raw material for the production of phosphoric acid, was supplied by prof. dr U. Glabisz (Polytechnic Institute, Szczecin). The material was sieved to get the particle size of $\varnothing < 0.125$ mm. The material was further homogenized by mixing for 16 hrs. in a plastic drum rotated in three directions. Preliminary homogeneity testing was performed determining Fe and Y contents by X-ray fluorescence in several subsamples of the material taken from the drum.

Distribution of 50 g portions into wide-mouthed, air-tight polyethylene bottles was achieved with the aid of a specially constructed teflon scoop. Final homogeneity testing was performed by instrumental neutron activation analysis, analyzing several subsamples taken from 6 individual bottles chosen at random and comparing the results with the results for 6 subsamples taken from one bottle.

Statistical comparison of results for Eu, La, Na, Nd, Tb and Th obtained in the two series of measurements confirmed good homogeneity of the material, at least for sample weight of ≥ 100 mg.

CERTIFICATION

Assigning of "recommended" (certified) values was done on the basis of world-wide intercomparison in which 58 laboratories participated. Their cooperation is gratefully acknowledged.

The laboratories contributed 3101 bits of chemical information (802 laboratory means) on 67 elements.

Statistical evaluation of results for individual elements was performed by the method published previously ¹⁾.

The outlying results from the population of laboratory averages were rejected by concurrent use of the four criteria i.e. those of Dixon, Grubbs, coefficient of skewness, and coefficient of kurtosis at significance level of 0.05, followed by calculation of overall mean, confidence limits etc.

The criteria used to decide whether the overall mean can be given the status of "recommended" (certified) value were as follows:

1. Relative uncertainty of the overall mean:

$$\frac{S.D. \cdot t_{0.05}}{\sqrt{N \cdot \bar{X}}} \leq 20\% \text{ (trace element)}$$

$$\leq 10\% \text{ (major elements)}$$

or relative standard deviation

$$\frac{S.D.}{\bar{X}} \leq 25\% \text{ (trace elements)}$$

$$\leq 15\% \text{ (major elements)}$$

For the purpose of this work elements with concentration exceeding 0.5 % (5000 ppm) were considered to be major elements.

2. The overall mean was calculated from at least 3 laboratory averages obtained by more than one analytical technique. If results from only one analytical technique are available the number of laboratory averages cannot be smaller than 5.

3. If the conditions (1) and (2) are fulfilled but the number of outliers exceeds 50 %, the additional procedure is activated which checks the changes of the mean and standard deviation respectively accompanying successive rejections.

The process of rejecting of outliers is then stopped when the change in both the mean and standard deviation becomes lower or equal to 15%.

"Information" values were assigned to those elements for which the results while not fulfilling conditions (1)-(3) still fulfilled the following condition:

$$\frac{S.D. \cdot t_{0.05}}{\sqrt{N \cdot \bar{X}}} \leq 50\% \text{ (trace element)}$$

$$\leq 30\% \text{ (major elements)}$$

"RECOMMENDED" VALUES FOR CTA-AC-1

MAJOR ELEMENTS

Element	Concentration [wt%]
Ca	32.7 ± 3.1
Si	0.57 ± 0.13

TRACE ELEMENTS

Element	Concentration [mg/kg (ppm)]	Element	Concentration [mg/kg (ppm)]
Ba	767 ± 79	Sc	0.244 ± 0.035
Ce	3326 ± 175	Sm	162 ± 24
Co	2.72 ± 0.28	Ta	2.65 ± 0.23
Cu	54.0 ± 4.5	Tb	13.9 ± 1.5
Eu	46.7 ± 1.3	Th	21.8 ± 2.1
Gd	124 ± 23	Ti	2927 ± 554
Hf	1.13 ± 0.17	U	4.4 ± 0.9
La	2176 ± 94	V	104 ± 10
Lu	1.08 ± 0.16	Y	272 ± 53
Mn	317 ± 50	Yb	11.4 ± 2.0
Na	3841 ± 564	Zn	38.0 ± 7.6
Nd	1087 ± 124		

"INFORMATION" VALUES FOR CTA-AC-1

MAJOR ELEMENTS

TRACE ELEMENTS

Element	Concentration [wt%]	Element	Concentration [mg/kg (ppm)]
Al	0.41	Cr	13
Fe	0.50	Dy	78
Sr	2.0	Er	26
		Ho	9
		K	2088
		Mg	435
		Ni	9
		Pr	353
		Zr	51

The preparation and certification of the CTA-AC-1 certified reference material was performed by the staff of the Department of Analytical Chemistry, Institute of Nuclear Chemistry and Technology, under the direction of prof. dr hab. R. Dybczyński

¹⁾ R. Dybczyński, *Analytica Chimica Acta*, **117**, 53-70 (1980)

LIST OF PARTICIPANTS OF THE INTERCOMPARISON CTA-AC-1 (in alphabetical order)

- Dr. I. Adamiec, Instytut Metali Nieżelaznych, Gliwice, POLAND
- Dr. F. Augustin, Bureau de Recherches Geologiques et Minieres, Orleans, FRANCE
- Dr. D. J. Bland, British Geological Survey, Nottingham, UNITED KINGDOM
- Dr. R. Bojanowski, Instytut Oceanologii PAN, Sopot, and Mr. A. Pietruszewski, CLOR, Warszawa, POLAND
- Mr. E. Bolibrzuch, Instytut Uprawy, Nawożenia i Gleboznawstwa, Puławy, POLAND
- Prof. dr. M. de Bruin, Interuniversity Reactor Institute, Delft, THE NETHERLANDS
- Mrs. T. Capała, Mrs. I. Sheybal, Instytut Metalurgii Żelaza, Gliwice, POLAND
- Prof. dr. Chai Chifang, Institute of High Energy Physics, Beijing, PEOPLE'S REPUBLIC OF CHINA
- Mr. W. Czajowski, Mrs. H. Szajbo, KGHM LUBIN, Lubin, POLAND
- Dr. M. Dermelj, "Jozef Stefan" Institute of Nuclear Chemistry, Ljubljana, YUGOSLAVIA
- Dr. P. Dulski, Hahn-Meitner Institute, WEST BERLIN
- Dr. M. Geisler, Dr. H. Schelhorn, Zentralinstitut für Isotopen-und Strahlenforschung, Leipzig, GERMAN DEMOCRATIC REPUBLIC
- Prof. dr. U. Glabisz, Politechnika Szczecińska, Szczecin POLAND
- Dr. K. Govindaraju, Centre de Recherches Petrographiques et Geochimiques, Vandoeuvre-les-Nancy, FRANCE
- Mr. W. Grysiewicz, Politechnika Wrocławska, Zakład Doświadczalny "Hydro-Mech" Kowary, POLAND
- Dr. C. Gunchew, Mrs. R. Georgiewa, Mr. W. Iwanow, Institute for Plant Protection, Kostinbrod near Sofia, BULGARIA
- Prof. J. Hertogen, Katholieke Universiteit Leuven, Leuven, BELGIUM
- Prof. dr. A. Hulanicki, Uniwersytet Warszawski, Warszawa, POLAND
- Mr. K. Jędrał, Mrs. T. Morawska, Mr. B. Lis, Państwowy Instytut Geologiczny, Warszawa POLAND
- Mr. Z. Jońca, Państwowy Instytut Geologiczny, Warszawa, POLAND
- Doc. dr. K. Kasiura, Politechnika Warszawska, Warszawa, POLAND
- Dr. L. Kinova, Institute of Nuclear Research and Nuclear Energy, Bulgarian Academy of Sciences, Sofia, BULGARIA

Dr. V. Kocman, Domtar Inc. Research Centre, Senneville Quebec, CANADA

Dr. G. M. Kolesov, V. I. Vernadsky, Institute of Geochemistry and Analytical Chemistry, Moscow, USSR

Dr. F. Kubanek, Hahn-Meitner Institute, WEST BERLIN

Dr. J. Kučera, Nuclear Research Institute, Rez near Prague, CZECHOSLOVAKIA

Dr. H. Kunzendorf, RISO National Laboratory, Roskilde, DENMARK

Dr. A. Lechotycki, Mr. J. Dudek, Institute of Nuclear Chemistry and Technology, Warszawa, POLAND

Dr. M. Lipponen, Technical Research Centre, of Finland Reactor Laboratory, Espoo, FINLAND

Mrs. M. Małusecka, "ENERGOPOMIAR", Katowice, POLAND

Dr. M. Michalewska, Politechnika Wrocławska, Wrocław, POLAND

Dr. K. Niebergall, Karl-Marx Universität, Leipzig, GERMAN DEMOCRATIC REPUBLIC

Dr. I. Obrusnik, Nuclear Research Institute, Rez near Prague, CZECHOSLOVAKIA

Prof. dr. N. N. Papadopoulos, Dr. S. Synetos, NRCPS Demokritos Technology, Attiki, and Dr. D. Tarenidis, Institute of Geological and Mineral Exploration, Athens, GREECE

Mr. M. Paul, Instytut Warzywnictwa, Skierniewice, POLAND

Dr. J. Pempkowiak, Instytut Oceanologii PAN, Sopot, POLAND

Dr. J. C. Philippot, Laboratoire de Metrologie de l'Environnement, Orsay, FRANCE

Dr. P. J. Potts, Open University, Milton Keynes, UNITED KINGDOM

Dr. A. Ramsza, Instytut Technologii Materiałów Elektronicznych, Warszawa, POLAND

Mr. E. J. Ring, MINTEK Analytical Services, Randburg, SOUTH AFRICA

Dr. I. Roelandts, Geology, University of Liege, Liege, BELGIUM

Dr. J. Samuel, Centre de Sedimentologie et de Geochimie de la Surface, Strasbourg, FRANCE

Dr. L. Savoyant, Dr. S. Pourtales, Centre Geologique et Geophysique, Montpellier FRANCE

Mr. A. Skalny, Ośrodek Badań i Kontroli Środowiska, Tarnów, POLAND

Dr. B. Skwarzec, Instytut Oceanologii PAN, Sopot, POLAND

Mrs. E. Sterlińska, Mrs. T. Kuras, Institute of Nuclear Chemistry and Technology, Warszawa, POLAND

Dr. M. Stępniewski, Dr. Paślawski, Państwowy Instytut Geologiczny, Warszawa, POLAND

Dr. W. Szacki, Hahn-Meitner Institute, WEST BERLIN

Mr. Z. Szałowski, Ośrodek Badań i Kontroli Środowiska, Legnica, POLAND

Mr. Z. Szopa, Dr. H. Polkowska-Motrenko, Prof. dr. R. Dybczyński, Miss. M. Mandecka, Mr. K. Kulisa, Mr. Z. Samczyński, Institute of Nuclear Chemistry and Technology, Warszawa, POLAND

Doc. dr. S. Szymczyk, Dr. E. Dutkiewicz, Instytut Fizyki Jądrowej, Kraków, POLAND

Mrs. H. Szyszko, Mrs. O. Grabowska, Mrs. E. Szybisz, Institute of Nuclear Chemistry and Technology, Warszawa, POLAND

Doc. dr. U. Tomza, Uniwersytet Śląski, Katowice, POLAND

Dr. M. Vobecky, Ustav Nuklearni Biologie a Radiochemie, Prague, CZECHOSLOVAKIA

Mrs. J. Wanat, Politechnika Wrocławska Zakład Doświadczalny "Hydro-Mech" Kowary, POLAND

Mr. L. Wawrzonek, Mrs. B. Małozewska-Bućko, Institute of Nuclear Chemistry and Technology, Warszawa, POLAND

Prof. Chen Yuanpan, Research Institute of Geology for Mineral Resources, Shanlidian, PEOPLE'S REPUBLIC OF CHINA

Mrs. L. Ząbczyńska, Mrs. K. Grzeszczuk, Zakłady Azotowe w Tarnowie, Tarnów, POLAND



**POLISH CERTIFIED REFERENCE MATERIAL
FOR MULTIELEMENT TRACE ANALYSIS**

APATITE CONCENTRATE (CTA-AC-1)

1. **The stability of Apatite Concentrate (CTA-AC-1) has been monitored during storage. On the basis of the obtained results, the shelf life is extended until 31 December 2032.**
2. The material is stored in an air conditioned room, in which the temperature does not exceed 22°C. Its stability is monitored at appropriate time intervals.
3. The users should store the material at room temperature $[(20\pm 5)^{\circ}\text{C}]$ in tightly closed original containers.
4. For chemical analysis and simultaneous determination of water content, samples of the material of adequate size should be taken. Immediately after sampling, it is necessary to cover firmly the container. The unused part of samples must not be placed back in the container.

Chajduk Ewelina, Phd

*Head of Laboratory
of Nuclear Analytical Methods*